

Interference by Drugs Contained in Over-the-counter Cold Syrups on Methamphetamine Immunoassay Test Kits Used in Drug Abuse Assessment

BIN-CHUNG HUANG, MEI-HWA LIEN, POR-YUH WANG AND BER-LIN CHANG

*National Laboratories of Foods and Drugs
Department of Health, Executive Yuan, Taiwan, R.O.C.*

ABSTRACT

Eight commercial methamphetamine test kits : AbuSign, accuPINCH, AccuSign, I.D. BLOCK, Medi-Mate, QuikPac II, SureStep and Visualine II were evaluated for their interferences by twelve drugs commonly found in over-the-counter cold syrups: aspirin, brompheniramine, chlorpheniramine, ephedrine, ethenzamide, guaifenesin, l-phenylephrine, methylephedrine, pheniramine, phentermine, phenylpropanolamine, and pseudoephedrine; and by clobenzorex, ranitidine which have been found to interfere in some amphetamine immunoassays. These fourteen drugs were evaluated by using measured concentrations ranging from 1 mg/L to 1000 mg/L. Results indicate that ephedrine and pseudoephedrine demonstrate a significant interference in most of test kits, while most analgesics, antihistamines and antitussives show no interference with the exception of brompheniramine. Furthermore, phenylpropanolamine, phentermine, phenylephrine, ranitidine and methylephedrine all show some degree of interference in Quikpac II, SureStep and Visualine II. While accuPINCH was shown to be the most specific test kit, SureStep and QuikPac II, which produced false positive results when tested at concentrations of 1 mg/L of ephedrine or pseudoephedrine, and thus demonstrate less specificity than all the other test kits.

Key words: Methamphetamine, test kit, interference, cold syrup.

INTRODUCTION

Cases of drug abuse with metamphetamine and heroin have significantly increased in Taiwan since 1991⁽¹⁾. Presently these two drugs are periodically tested in both foreign laborers and drug addicts. Amphetamine-like compounds, such as phenylpropanolamine, ephedrine, pseudoephedrine, which often appear in over-the-counter(OTC) cold syrups, may produce false positive results in some commercial immunoas-

say systems for amphetamines. Therefore, it is necessary for a laboratory to have a sound, properly validated confirmation assay to identify the drugs of interest. The Mandatory Guidelines for Federal Workplace Drug Testing⁽²⁾ recommended by the Department of Health and Human Service(DHHS) in U.S.A. regulate an initial test that is based on an immunoassay to eliminate negative urine specimens from further consideration and a confirmatory test that requires the use of a gas chromatography/mass spectrometry

Table 1. Specificity data of commonly used sympathomimetics in OTC cold syrups on package insert of different test kits

Unit:mg / L

Test Kit	Ephedrine	Pseudo-ephedrine	Phenylpropanolamine	Methylephedrine	Phenylephrine	Phentermine
AbuSign	+ / 90	+ / 60	-	*	-	-
accuPINCH	- / 100	- / 100	- / 100	- / 100	*	- / 100
AccuSign	*	*	-	*	-	-
I.D. BLOCK	+ / 10	*	*	*	*	+ / 10
Medi-Mate	+ / 5	+ / 20	-	*	-	-
QuikPac II	*	*	*	*	*	*
SureStep	*	*	*	*	*	*
Visualine II	+ / 100	+ / 10	- / 100	*	*	*

Notes: + / 100 : produce interference when tested at concentration of 100 mg / L
 - / 100 : do not produce interference when tested at concentration of 100 mg / L
 - : produce interference when tested (did not report its concentration)
 * : No specificity data on package insert.

(GC/MS) method if a specimen is identified as positive on the initial test. In our country, the guidelines of drug testing also follow the similar regulations as those in U.S.A.

Recently, problems have been reported in the analysis of methamphetamine. Specifically, several specimens have been reported by drug testing laboratories in Taiwan to be positive for methamphetamine which were subsequently confirmed to be false positive. These specimens have several common characteristics, including not only the absence of methamphetamine but also the presence of extremely high concentrations of ephedrine or pseudoephedrine⁽³⁾. These problems had previously appeared in the United States, and lead the DHHS to require an amphetamine concentration greater than 200 mg/L in addition to a methamphetamine concentration greater than 500 mg/L before a sample is reported as positive for

methamphetamine⁽²⁾. Because the general populace of Taiwan are accustomed to buying and self-administering OTC cold syrups from drug stores whenever they contract a cold, the problem mentioned above seem more formidable in Taiwan than in the United States.

Generally there are sixteen drugs commonly found in OTC cold syrups⁽⁴⁾ including acetaminophen, aspirin, brompheniramine, chlorpheniramine, dextromethorphan, codeine, ephedrine, ethenzamide, guaifenesin, ibuprofen, l-phenylephrine, methylephedrine, pheniramine, phentermine, phenylpropanolamine, and pseudoephedrine. These drugs can be categorized into four classes: sympathomimetics : ephedrine, pseudoephedrine, phenylpropanolamine, methylephedrine, phenylephrine, phentermine; analgesics : acetaminophen, ibuprofen, ethenzamide, aspirin; antihistamines : chlorpheniramine,

Journal of Food and Drug Analysis, 1995, 3(4)

brompheniramine, pheniramine; antitussives : dextromethorphan, codeine, guaifenesin. Two other drugs : clobenzorex, ranitidine, which are not routinely added to cold syrups have also been reported to interfere some immunoassays^(5,6). Therefore, a total of eighteen drugs were assessed in the present study.

Specificity data of commonly used drugs are often reported on the package insert of each test kit⁽⁷⁻¹⁴⁾ and are summarized in Table 1~5. Table 1 lists the specificity data for the sympathomimetics; Table 2 for the analgesics; Table 3 for the antihistamines; Table 4 for the antitussives; Table 5 for clobenzorex and ranitidine. Table 1

Table 2. Specificity insert of commonly used analgesics in OTC cold syrups on package insert of different test kits unit:mg / L

Test kit	Acetaminophen	Ibuprofen	Ethenzamide	Aspirin
AbuSign	—	—	—	—
accuPINCH	— / 100	— / 100	*	*
AccuSign	—	*	*	*
I.D. BLOCK	*	*	*	*
Medi-Mate	—	*	*	*
QuikPac II	*	*	*	*
SureStep	*	*	*	*
Visualine II	— / 100	— / 100	*	*

Note: — / 100 : do not produce interference when tested at concentration of 100 mg / L
 — : produce interference when tested (did not report its concentration)
 * : No specificity data on package insert.

Table 3. Specificity insert of commonly used antihistamines in OTC cold syrups on package insert of different test kits unit:mg / L

Test Kit	Chlorpheniramine	Brompheniramine	pheniramine
AbuSign	—	*	*
accuPINCH	*	*	*
AccuSign	—	*	*
I.D. BLOCK	*	*	*
Medi-Mate	—	*	*
QuikPac II	*	*	*
SureStep	*	*	*
Visualine II	*	— / 100	*

Note: — / 100 : do not produce interference when tested at concentration of 100 mg / L
 — : produce interference when tested (did not report its concentration)
 * : No specificity data on package insert.

Journal of Food and Drug Analysis, 1995, 3(4)

Table 4. Specificity insert of commonly used anti-tussives in OTC cold syrups on package insert of different test kits

Test Kit	Dextromethorphan	Codeine	Guaifenesin
AbuSign	—	—	*
accuPINCH	*	— / 100	*
AccSign	—	—	*
I.D. BLOCK	*	*	*
Medi-Mate	—	—	*
QuikPac II	*	*	*
SureStep	*	*	*
Visualine II	*	*	*

Note: — / 100 : do not produce interference when tested at concentration of 100 mg / L
 — : produce interference when tested (did not report its concentration)
 * : No specificity data on package insert.

shows that ephedrine and pseudoephedrine are apt to produce cross reactions, and other drugs either show no cross reaction, or do not have their specificity data. Because some package inserts, such as those of QuikPac II, SureStep, do not contain specificity data related to these drugs, and no test kit has complete specificity data for these sympathomimetics, it is necessary to evaluate the overall specificity data of these sympathomimetics.

Aside from sympathomimetics, other drugs, such as analgesics(Table 2), antihistamines(Table 3), antitussives(Table 4), either show little or no specificity data or do not produce interference according to their package inserts. Table 2 and Table 4 show that acetaminophen, ibuprofen, dextromethorphan, and codeine did not produce any significant interference. Therefore, these four drugs were not included in this study.

Because of variability in commercial

Table 5. Specificity insert of two cross-reactive drugs on package insert of different test kits

Test Kit	Clobenzorex	Ranitidine
AbuSign	*	—
accuPINCH	*	— / 100
AccSign	*	—
I.D. BLOCK	*	*
Medi-Mate	*	—
QuikPac II	*	*
SureStep	*	*
Visualine II	*	*

Note: + / 100 : produce interference when tested at concentration of 100 mg/L
 — / 100 : do not produce interference when tested at concentration of 100 mg / L
 — : produce interference when tested (did not report its concentration)
 * : No specificity data on package insert.

immunoassay kits, a comprehensive and periodic evaluation of immunoassay characteristics is required for selecting an appropriate initial test for the drug of abuse. The cross-reactivity of amphetamine analogs or other interfering compounds have already been evaluated in several immunoassays⁽¹⁵⁻²⁴⁾. However, most reports evaluated either automatic analyzers only, such as EMIT⁽¹⁵⁻¹⁶⁾, TDX⁽¹⁷⁾, their method comparison⁽¹⁸⁻²⁰⁾, or the immunoassay test kits from the same manufacturers⁽²¹⁻²⁴⁾. Furthermore these papers mainly focused on amphetamine analogs only, and they did not pay much attention to the other compounds in OTC syrups. In 1992, Jeffrey et al.⁽²⁰⁾ examined six commercial immunoassays for their cross-reactivity to ephedrine, pseudoephedrine, and phenylpropanolamine, and their results showed that all specimens spiked at 100

Table 6. Test results of sympathomimetics commonly used in OTC cold syrups on eight test kits

Unit:mg / L

Test Kit	Ephedrine	Pseudo-ephedrine	Phenylpropanolamine	Methylephedrine	Phenylephrine	Phentermine
AbuSign	± / 10	± / 100	- / 1000	- / 1000	± / 10	- / 1000
accuPINCH	- / 1000	± / 100	- / 1000	- / 1000	- / 1000	- / 1000
AccuSign	+ / 50	+ / 500	- / 1000	- / 1000	- / 1000	- / 1000
I.D. BLOCK	+ / 100	± / 50	± / 500	- / 1000	± / 10	± / 500
Medi-Mate	+ / 10	± / 50	- / 1000	- / 1000	+ / 500	- / 1000
QuikPac II	+ / 1	+ / 1	± / 100	± / 500	+ / 50	± / 100
SureStep	+ / 1	+ / 1	± / 500	± / 500	+ / 100	+ / 500
Visualine II	+ / 100	+ / 10	± / 1000	± / 1000	± / 1000	+ / 1000

Notes: + / 50 : response greater than or equal to 1 mg / L when tested at concentration of 50 mg / L
 - / 1000 : response less than 1 mg / L when tested at concentration of 1000 mg / L
 ± / 1000 :response near 1mg / L cutoff when tested at concentration of 1000 mg / L

mg/L were negative as judged by all immunoassays tested. Because these drugs are sold in some OTC cold preparations and are excreted into urine largely as the parent drugs, it is possible that their concentrations in the range of 1000 mg/L might be encountered in urine specimens, and thus evaluation of higher concentrations is also necessary. Therefore, concentrations of drugs in OTC cold syrups ranging from 1 mg/L to 1000 mg/L in urine were included in the study.

MATERIALS AND METHODS

I. Materials

Aspirin, brompheniramine maleate, chlorpheniramine maleate, ephedrine HCl, guaifenesin, methamphetamine HCl, l-phenylephrine HCl, methylephedrine, pheniramine maleate, phentermine, phenylpropanolamine HCl, pseudoephedrine HCl and ranitidine HCl were purchased from Sigma Chemical Co.. Ethenzamide were purchased from Aldrich Co.. Clobenzorex is a reference standard of our laboratory. All stan-

dards were reagent grade.

II. Immunoassay Kits

HYCOR accuPINCH™ Methamphetamine Test test kits⁽⁷⁾ (Lot. No. 35178C3) were obtained from Hycor Biomedical Inc., Methamphetamine I.D. BLOCK™ Detection Kit test kits⁽⁸⁾ (Lot. No. 13994-MAB-1) were obtained from International Diagnostic Systems, AbuSign™ MET One-Step Methamphetamine Test test kits⁽⁹⁾ (Lot. No.856401) were obtained from Princeton BioMeditech, AccuSign™ MET New One-Step Methamphetamine Test test kits⁽¹⁰⁾ (Lot. No. 858101) were obtained from Bio AccuTech Inc., Medi-Mate™ MET One Step Methamphetamine Test test kits⁽¹¹⁾ (Lot. No. 857201) were obtained from Medi-Mate Technologies Inc., QuikPac II - One Step Urine Amphetamine Test test kits⁽¹²⁾ (Lot. No. 010743) were obtained from Syntro BioResearch, Inc. SureStep™ Drug Screen Methamphetamine test kits⁽¹³⁾ (Lot. No. 201) were obtained from Applied Biotech, Inc., Visualine II™ Methamphetamine Test One-Step Urine Drug Screening test kits⁽¹⁴⁾ (Lot. No.

Table 7. Test results of analgesics, antihistamines, and antitussives commonly used in OTC cold syrups on eight test kits
unit: mg / L

Test Kit	Ethenzamide	Aspirin	Chlorpheniramine	Brompheniramine	Pheniramine	Guaifenesin
AbuSign	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000
accuPINCH	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000
AccuSign	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000
I.D. BLOCK	- / 1000	- / 1000	- / 1000	+ / 1000	- / 1000	- / 1000
Medi-Mate	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000
QuikPac II	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000
SureStep	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000
Visualine II	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000	- / 1000

Notes: + / 1000: response greater than or equal to 1 mg / L when tested at concentration of 1000 mg / L
- / 1000: response less than 1 mg / L when tested at concentration of 1000 mg / L

82509) were obtained from Hanson Hong Biomedical Co. Ltd.

III. Control and Standard Urines

Methamphetamine control urines, Amphetamines positive specimen (Lot. No.57702) and negative specimen (Lot. No. 57711), were purchased from BIO-RAD Laboratories. Methamphetamine standard urine at concentration of 1 mg/L cutoff was prepared from spiking methamphetamine HCl in aforementioned negative specimen.

IV. Preparations of Drugs in Urine

The test solutions of aspirin, brompheniramine, chlorpheniramine, clobenzorex, ephedrine, ethenzamide, guaifenesin, l-phenylephrine, methylephedrine, pheniramine, phentermine, phenylpropanolamine, pseudoephedrine and ranitidine were prepared in a serial concentrations ranging from 1000, 500, 100, 50, 10, to 1 mg/L, if necessary, by spiking each of these drugs in negative specimen respectively.

V. Procedure

Each drug was tested in duplicate for every test kit by the following procedure. Firstly, each drug was tested at concentration of 1000 mg/L. Then we either continued to evaluate the test kits at lower concentrations in sequence of 500, 100, 50, 10, 1 mg/L if their results were positive, which mean the response was greater than or equal to that of 1 mg/L cutoff, or until the result was negative and then recorded the previous concentration, which means the response was less than that of 1 mg/L cutoff.

RESULTS

I. Sympathomimetic Amines

The results of interference by sympathomimetics commonly used in cold syrups in eight methamphetamine immunoassay test kits are listed in Table 6. The table clearly indicates that many sympathomimetics are apt to produce cross reactions in these immunoassay test kits at specified concentrations. Pseudoephedrine is the most

Table 8. Test results of clobenzorex, ranitidine on eight test kits

unit:mg / L

Test kit	Clobenzorex	Ranitidine
AbuSign	- / 1000	- / 1000
accuPINCH	- / 1000	- / 1000
AccuSign	- / 1000	- / 1000
I.D. BLOCK	± / 1000	+ / 1000
Medi-Mate	- / 1000	- / 1000
QuikPac II	- / 1000	± / 50
SureStep	- / 1000	± / 50
Visualine II	- / 1000	± / 1000

Notes: + / 1000 : response greater than or equal to 1 mg / L when tested at concentration of 1000 mg / L

- / 1000 : response less than 1 mg / L when tested at concentration of 1000 mg / L

± / 1000 : response near 1 mg / L cutoff when tested at concentration of 1000 mg / L

cross-reactive drug producing positive response even at a concentration of 1 mg/L in the SureStep test kit and the QuikPac II test kit. Ephedrine also produces significant interference in these test kits with the exception of the accuPINCH test kit. Phenylephrine also demonstrates little interference in test kits with the exception of the accuPINCH and AccuSign test kit. In addition, phenylpropanolamine produces little interference in the Visualine II, SureStep, QuikPac II, and the I.D.Block test kits; Phentermine produces little interference in the Visualine II, SureStep, QuikPac II, and the I.D.Block test kits. Methylephedrine also demonstrates little interference in the Visualine II, SureStep, and the QuikPac II test kits. The latter drug is the least cross-reactive sympathomimetic that only produces false positive results at concentrations above 500 mg/L. Table 6 clearly shows that ephedrine and pseudoephedrine are more cross-reactive than all the other sympathomimetics.

II. Analgesics, Antihistamines, and Antitussives

The results of interference by the analgesics, antihistamines, and antitussives commonly used

in cold syrups in eight methamphetamine immunoassay test kits are listed in Table 7. The table clearly indicates that most of these drugs do not effect any significant levels of interference in these test kits except brompheniramine in the I.D. block test kit at 1000 mg/L.

III. Clobenzorex, Ranitidine

The results of interference by clobenzorex, ranitidine in eight methamphetamine immunoassay test kits are listed in Table 8. The table clearly indicates that clobenzorex does not produce a significant interference except for the I.D. block test kit, while ranitidine produces little interference in the I.D. block, QuikPac II, SureStep, and the Visualine II test kit, of which the QuikPac II and the SureStep test kits produce false positives at about 50 mg/L of ranitidine.

DISCUSSION

Cross-reactivity patterns with the sympathomimetic amines, clobenzorex, and ranitidine for methamphetamine test kits differed among these immunoassays. The sympathomimetic amines,

Journal of Food and Drug Analysis. 1995. 3(4)

especially pseudoephedrine and ephedrine, are more cross-reactive than all the other drug classes. Clobenzorex and ranitidine also exhibits a degree of interference on some of these test kits, but the analgesics, the antihistamines, and the antitussives do not effect any significant interference except brompheniramine in the I.D. block test kit at 1000 mg/L.

This study demonstrates that the QuikPac II and SureStep test kits have a greater chance of producing false positive amphetamine screening results due to the presence of a high concentration of the sympathomimetic amines, such as pseudoephedrine, ephedrine, phenylpropanolamine, phentermine, methylephedrine, and phenylephedrine in OTC cold syrups. These sympathomimetic amines may also produce false positive results for other test kits, including the AbuSign, AccuSign, I.D. block, MediMate, and the Visualine II test kits. In contrast, the accuPINCH test kit, which produces false positive result only at 100 mg/L of pseudoephedrine, is the most specific among these test kits.

Because some drugs in OTC cold syrups can cause interference in methamphetamine immunoassay test kits, a urine drug testing laboratory should establish a routine procedure that includes at least one immunoassay served as the initial screening method and a second nonimmunoassay method which confirmed the putative positive result from the initial test. Information from this study may be of use in selecting a method to ensure that the combination of the two methods does not produce a false positive result in routine work.

ACKNOWLEDGEMENTS

This work was supported by National Laboratories of Foods and Drugs, Department of Health, Executive Yuan, R.O.C.

REFERENCES

1. Chien C.S., Huang B.C., Huang W.F. 1993. Testing of Drug Abuse in Taiwan Area. pp.17-20, International Symposium on Drug Abuse, Taiwan, R.O.C.
2. Mandatory Guidelines for Federal Workplace Drug Testing Programs; Notice. 9 June 1994. Federal Register, 59(110):29908-29931.
3. Lin W.L., Deng J.F. 1993. Drug Testing in Clinical Toxicology Lab VGH, pp.81-100, International Symposium on Drug Abuse, Taiwan, R.O.C.
4. 蔡清彥. 1991. Preparations for Common Cold, pp.284-294, Handbook of Common Drugs, 玉山書局。
5. John A.T. 1994. Amphetamine-Positive Drug Screens from Use of Clobenzorex Hydrochloride. J. of Analytic Toxicology, 18:183
6. Alphone P. and Karen V.H. and Jay S. 1991. Ranitidine Interference with the Monoclonal EMIT d.a.u. Amphetamine/Methamphetamine Immunoassay. J. of Analytic Toxicology, 15:101~103.
7. accuPINCH™ Methamphetamine TEST, An Enzyme Immunoassay for Methamphetamine In Human Urine. 1994. Package Insert. Hycor Biomedical Inc.
8. I.D. BLOCK™ Methamphetamine Detection Kit. 1994. Package Insert. International Diagnostic Systems Corp.
9. AbuSign™ MET New One-Step Methamphetamine Test. 1994. Package Insert. Princeton BioMeditech Corporation.
10. AccuSign™ MET New One-Step Methamphetamine Test. 1994. Package Insert. Bio AccuTech Inc.
11. Medi-Mate™ MET One Step Methamphetamine Test. 1994. Package Insert. Medi-Mate Technologies Inc.
12. QuikPac II™ One Step Urine Amphetamine Test. 1994. Package Insert. Syntron Bioresearch Inc.
13. SureStep™ Drug Screen Methamphetamine. 1994 : Package Insert. Applied Biotech, Inc.
14. Visualine II™ Methamphetamine Test One-Step Urine Drug Screening. 1994. Package

Journal of Food and Drug Analysis. 1995. 3(4)

- Insert. Hanson Hong Biomedical Co. Ltd.
15. Robert D.B. 1981. Amphetamine EMIT Structure Versus Reactivity. *Clin. Toxicol.* 18:91-110.
 16. Gregory F.F. 1989. Ranitidine and High Concentration of Phenylpropanolamine Cross React in the EMIT Monoclonal Amphetamine /Methamphetamine Immunoassay, *Clin. Chem.* 35:1998-1999.
 17. Cody T.C., and Robert S. 1993. Fluorescence Polarization Immunoassay Detection of Amphetamine, Methamphetamine, and Illicit Amphetamine Analogues, *J. of Anal. Toxicol.* 17:26-30.
 18. Prezekop M.A., Manno J.E., Kunsman G.W., Cockerham K.R., and Manno B.R. 1991. Evaluation of the Abbott ADx Amphetamine/ Methamphetamine II Abused Drug Assay: Comparison to TDx, Emit, and GC/MS Methods. *J. of Anal. Toxicol.* 15:323-326.
 19. Werawan R. and David E.M. 1991. Comparison of Three Commercial Amphetamine Immunoassays for Detection of Methamphetamine, Methylenedioxy- amphetamine, Methylenedioxymethamphetamine, and Methylenedioxyethylamphetamine, *J. of Anal. Toxicol.* 12:229-233.
 20. Jeffrey D., Jones R., Barry L., and Michael L.S. 1992. Evaluation of Six Commercial Amphetamine and Methamphetamine Immunoassays for Cross-reactivity to Phenylpropanolamine and Ephedrine in Urine. *J. of Anal. Toxicol.* 16:211-213.
 21. John T.C.. 1990. Detection of d,l-Amphetamine, d,l-Methamphetamine, and Illicit Amphetamine Analogs Using Diagnostic Products Corporation's Amphetamine and Methamphetamine Radioimmunoassay. *J. of Anal. Toxicol.* 14:321-324.
 22. Robinson C.A., Morrison M., and Ketchum C.H. 1990. Evaluation of on Site THC Screening Products. *Clin. Chem.* 36:1027.
 23. Alan H.B. Wu, Kent J., James C., Shu D.X., and Shan S. Wang. 1992. Sensitivity and Specificity of the Triage TM Panel for Screening Abuse of Phencyclidine, Amphetamines, Opiates, and Tetrahydrocannabinol in Urine. *Clin. Chem.* 38:1007-1008.
 24. John T.C.. 1990. Cross-reactivity of Amphetamine Analogues with Roche Abuscreen Radioimmunoassay Reagents. *J. of Anal. Toxicol.* 14:50-53.

國內常見感冒糖漿劑成分對尿液中甲基安非他命篩檢之影響

黃炳中 連美華 王博譽 張柏林

行政院衛生署藥物食品檢驗局

摘 要

本計畫係針對感冒製劑成分，對尿液中甲基安非他命在免疫學分析套組試驗中所可能產生的交叉反應進行評估，以期瞭解各套組之特異性，供作選用尿液檢體初步篩選用檢測套組之參考。

本計畫評估之免疫學分析套組有：AbuSign, accuPINCH, AccuSign, I.D. BLOCK, Medi-Mate, QuikPac II, SureStep及Visualine II等八種，列入評估之藥物有：aspirin, brompheniramine, chlorpheniramine, clobenzorex, ephedrine, ethenzamide, guaifenesin, l-phenylephrine, methylephedrine, pheniramine, phentermine, phenylpropanolamine, pseudoephedrine及ranitidine等十四種。

結果顯示，ephedrine及pseudoephedrine產生交叉反應之濃度最低，即上述藥物於 $1 \mu\text{g} / \text{ml}$ 時，以QuikPac II及SureStep檢測即呈偽陽性反應。若以藥物濃度為 $1000 \mu\text{g} / \text{ml}$ 時，仍

不呈陽性或疑似陽性反應為判定交叉反應之界限，則pseudoephedrine對八種套組均呈交叉反應。ephedrine對於AccuPINCH外之套組，均呈交叉反應。Phenylephrine對除accuPINCH, Accusign外之套組，均呈交叉反應。Phentermine及phenylpropanolamine對除Abusign, AccuPINCH, Accusign, Medi-Mate外之套組，均呈交叉反應。phentermine及phenylpropanolamine對除Abusign, AccuPINCH, Accusign, Medi-Mate外之套組，均呈交叉反應。Methylephedrine對QuikPac II, SureStep及Visualine II套組呈交叉反應，對其他不呈交叉反應。此外ethenzamide, aspirin, chlorpheniramine, pheniramine及guaifenesin對八種套組均不呈交叉反應；而brompheniramine及clobenzorex只對I.D. Block套組產生交叉反應；ranitidine對除Abusign, AccuPINCH, Accusign及Medi-Mate外之套組均呈交叉反應。